

Protocol# 2 - Wipe Tests.lsa

User: Default

Assay Definition

Assay Description:

Assay Type: DPM (Single)

Report Name: Report1

Output Data Path: C:\Packard\Tricarb\Results\Default\Wipe Tests\20200202_1121

Raw Results Path: C:\Packard\Tricarb\Results\Default\Wipe Tests\20200202_1121\20200202_1121.results

Comma-Delimited File Name: C:\Packard\Tricarb\Results\Default\Wipe Tests\20200202_1121\Report1.txt

Assay File Name: C:\Packard\TriCarb\Assays\Wipe Tests.lsa

Count Conditions

Nuclide: 14C

Quench Indicator: tSIE/AEC

External Std Terminator (sec): 0.5 2s%

Pre-Count Delay (min): 0.00

Quench Set:

Low Energy: 14C

Count Time (min): 1.00

Count Mode: Normal

Assay Count Cycles: 1

Repeat Sample Count: 1

Number of Vials/Sample: 1

Calculate % Reference: Off

Background Subtract

Background Subtract: Off

Low CPM Threshold: Off

2 Sigma % Terminator: Off

Regions	LL	UL
A	0.0	156.0
B	4.0	156.0
C	0.0	0.0

Count Corrections

Static Controller: On

Luminescence Correction: Off

GCT: n/a

Colored Samples: Off

Heterogeneity Monitor: Off

PAC: n/a

Coincidence Time (nsec): 18

Delay Before Burst (nsec): 75

PAC Strength: n/a

Auxiliary Spectrum: n/a

Cycle 1 Results

S#	Count	Time	CPMA	DPM1	SIS	tSIE	MESSAGES
1	1.00	44	47	539.59	347.10		
2	1.00	29	31	465.55	330.07		
3	1.00	62	73	101.49	142.71		
4	1.00	29	31	575.36	325.14		
5	1.00	34	36	637.72	331.11		
6	1.00	24	28	417.58	174.05		
7	1.00	25	27	662.20	327.25		
8	1.00	25	27	637.12	339.44		
9	1.00	20	22	398.09	198.58		
10	1.00	27	30	599.57	213.00		
11	1.00	20	22	368.49	193.60		

Protocol# 2 - Wipe Tests.lsa

User: Default

12	1.00	29	31	464.73	301.13
13	1.00	32	34	806.48	454.76