

Biospherical Instruments Inc

CALIBRATION CERTIFICATE

UNDERWATER PAR SENSOR WITH LOG AMPLIFIER

Calibration Date: 05/04/10

Model Number: QSP-200L

Serial Number: 4479

Operator: TPC

Standard Lamp: GS1019(8/28/08)

Job No.: R-10610

Operating Voltage Range: 6 to 15 VDC (+)

Note: The QSP-200 uses a log amplifier to measure the detector signal current with $V = \log I (\text{Amps}) / I_{\text{Ref}}$
To calculate irradiance, use this formula:

$$\text{Irradiance} = \text{Calibration factor} * (10^{\text{Light Signal Voltage}} - 10^{\text{Dark Voltage}})$$

With the appropriate (solar corrected) Irradiance Calibration Factor:

Dry Calibration Factor:	9.15E+12	quanta/cm ² ·sec/"amps"	1.52E-05	μEinsteins/cm ² ·sec/"amps"
Wet Calibration Factor:	1.54E+13	quanta/cm ² ·sec/"amps"	2.56E-05	μEinsteins/cm ² ·sec/"amps"

Sensor Test Data and Results⁴⁾

Sensor Supply Current (Dark):		73.0	mA							
Supply Voltage:		6	Volts							
Lamp Integrated PAR Irradiance:		9.09E+15	quanta/cm ² ·sec	0.01509	μEinsteins/cm ² ·sec					
SC3 Immersion Coefficient:		0.594	Scalar Correction:	1	PAR Solar Correction:					1.0000
Nominal	Calibrated	Sensor	Measured	Measured	Estimated	Calc.				Test Irrad.
Filter OD	Trans.	Voltage	Trans.	Signal	Signal	Output	Error			(quanta/
				(Amps)	(Amps)	(Volts)	(Volts)	Error (%)		cm ² ·sec)
No Filter	100.00%	2.998	100.00%	9.95E-08	9.95E-08	2.999	0.001	0.0		9.09E+15
0.3	36.10%	2.558	36.19%	3.60E-08	3.59E-08	2.558	0.000	-0.3		3.29E+15
0.5	27.60%	2.445	27.83%	2.77E-08	2.75E-08	2.442	-0.003	-0.8		2.53E+15
1	9.27%	1.981	9.44%	9.40E-09	9.23E-09	1.973	-0.007	-1.8		8.58E+14
2	1.11%	1.117	1.14%	1.13E-09	1.10E-09	1.109	-0.008	-2.3		1.03E+14
3	0.05%	0.406	0.08%	7.58E-11	5.32E-11	0.366	-0.040	-29.9		6.92E+12

Dark Before: 0.252 Volts

Light - No Filter Hldr.: 2.998 Volts

Dark After - NFH: 0.254 Volts

Average Dark 0.253 Volts

$I_{\text{Ref}} = 1.00\text{E-}10$ Amps

$I_{\text{Dark}} = 1.79\text{E-}10$ Amps

$10^{V_{\text{Dark}}} = 1.789575$ Amps

RG780

0.3204

Notes:

1. Annual calibration is recommended.
2. There is increasing error associated with readings below zero.
3. The collector should be cleaned frequently with alcohol.
- 4) This section is for internal use and for more advanced analysis.