

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1431
CALIBRATION DATE: 25-Jul-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

GHIJ COEFFICIENTS

g = -4.10091904e+000
h = 5.32229305e-001
i = -1.62959890e-004
j = 3.69287371e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.23574371e-005
b = 5.31800108e-001
c = -4.09999848e+000
d = -8.87761066e-005
m = 4.3
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.77626	0.00000	0.00000
-0.9999	34.8839	2.80952	7.77067	2.80951	-0.00001
1.0001	34.8850	2.98129	7.97466	2.98129	0.00000
15.0001	34.8864	4.27931	9.37174	4.27931	0.00000
18.5001	34.8861	4.62664	9.71099	4.62665	0.00001
29.0000	34.8855	5.71237	10.70090	5.71234	-0.00003
32.5000	34.8790	6.08568	11.02031	6.08571	0.00002


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



17-Sep-10 0.9993910
25-Jul-12 1.0000000