

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1714  
CALIBRATION DATE: 25-Sep-08

SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

g = 4.79486569e-003  
h = 6.58260474e-004  
i = 2.03239098e-005  
j = 1.34593667e-006  
f0 = 1000.0

## IPTS-68 COEFFICIENTS

a = 3.68120702e-003  
b = 5.98900313e-004  
c = 1.31715847e-005  
d = 1.34714501e-006  
f0 = 5917.622


| BATH TEMP<br>(ITS-90) | INSTRUMENT FREQ<br>(Hz) | INST TEMP<br>(ITS-90) | RESIDUAL<br>(ITS-90) |
|-----------------------|-------------------------|-----------------------|----------------------|
| -1.5002               | 5917.622                | -1.4996               | 0.00059              |
| 0.9998                | 6259.148                | 0.9993                | -0.00050             |
| 4.4998                | 6760.852                | 4.4990                | -0.00078             |
| 7.9998                | 7290.703                | 7.9998                | -0.00001             |
| 11.4998               | 7849.288                | 11.5005               | 0.00074              |
| 14.9998               | 8437.173                | 15.0004               | 0.00059              |
| 18.4998               | 9055.120                | 18.4998               | 0.00002              |
| 21.9999               | 9703.915                | 21.9995               | -0.00043             |
| 25.4998               | 10384.218               | 25.4994               | -0.00044             |
| 28.9998               | 11096.696               | 28.9996               | -0.00019             |
| 32.4998               | 11841.958               | 32.5002               | 0.00041              |

Temperature ITS-90 =  $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$  (°C)

Temperature IPTS-68 =  $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature





# SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Temperature Calibration Report

|              |                                     |                 |           |
|--------------|-------------------------------------|-----------------|-----------|
| Customer:    | Woods Hole Oceangraphic Institution |                 |           |
| Job Number:  | 51761                               | Date of Report: | 9/25/2008 |
| Model Number | SBE 03-02/F                         | Serial Number:  | 031714    |

*Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.*

*An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.*

### 'AS RECEIVED CALIBRATION'

☒ Performed ☐ Not Performed

Date: 9/25/2008

Drift since last cal: +0.00003 Degrees Celsius/year

Comments:

### 'CALIBRATION AFTER REPAIR'

☐ Performed ☒ Not Performed

Date:

Drift since Last cal: Degrees Celsius/year

Comments: