

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration

Date April 30, 2008 S/N# CST-1117DR Pathlength 25 cm

Analog meter

V_d 0.063 V
 V_{air} 4.774 V
 V_{ref} 4.671 V

Temperature of calibration water 21.1 °C
Ambient temperature during calibration 25.1 °C

Relationship of transmittance (T_r) to beam attenuation coefficient (c), and pathlength (x): $T_r = e^{-cx}$

To determine beam transmittance: $T_r = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(T_r)$

V_d Meter output with the beam blocked. This is the offset.
 V_{air} Meter output in air with a clear beam path.
 V_{ref} Meter output with clean water in the path.
Temperature of calibration water: temperature of clean water used to obtain V_{ref} .
Ambient temperature: meter temperature in air during the calibration.
 V_{sig} Measured signal output of meter.