

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2880
CALIBRATION DATE: 12-Jan-10

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.06602669e+001
h = 1.44142660e+000
i = 2.34005419e-004
j = 5.73830122e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.88559389e-004
b = 1.44156294e+000
c = -1.06603497e+001
d = -8.38391225e-005
m = 3.6
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.71849	0.00000	0.00000
-1.0001	34.7618	2.80059	5.17433	2.80057	-0.00002
0.9999	34.7616	2.97173	5.28747	2.97175	0.00003
14.9999	34.7624	4.26569	6.07453	4.26568	-0.00001
18.4999	34.7622	4.61196	6.26832	4.61194	-0.00001
28.9999	34.7604	5.69418	6.83846	5.69423	0.00004
32.4999	34.7564	6.06671	7.02389	6.06669	-0.00003

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

