

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## C-Star Calibration

Date April 30, 2008 S/N# CST-1118DR Pathlength 25 cm

### Analog meter

$V_d$  0.061 V  
 $V_{air}$  4.826 V  
 $V_{ref}$  4.728 V

Temperature of calibration water 20.5 °C  
Ambient temperature during calibration 25.4 °C

Relationship of transmittance ( $T_r$ ) to beam attenuation coefficient ( $c$ ), and pathlength ( $x$ ):  $T_r = e^{-cx}$

To determine beam transmittance:  $T_r = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient:  $c = -1/x * \ln(T_r)$

$V_d$  Meter output with the beam blocked. This is the offset.

$V_{air}$  Meter output in air with a clear beam path.

$V_{ref}$  Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain  $V_{ref}$ .

Ambient temperature: meter temperature in air during the calibration.

$V_{sig}$  Measured signal output of meter.