



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0113  
 CALIBRATION DATE: 09-May-17

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS:            A = -3.4961e-003  
 Soc = 0.4615                B = 1.5746e-004  
 Voffset = -0.5045        C = -2.3216e-006  
 Tau20 = 1.00              E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS  
 D1 = 1.92634e-4        H1 = -3.300000e-2  
 D2 = -4.64803e-2      H2 = 5.00000e+3  
 H3 = 1.45000e+3

BATH OXYGEN (ml/l)	BATH TEMPERATURE (° C)	BATH SALINITY (PSU)	INSTRUMENT OUTPUT (volts)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.07	2.00	0.00	0.745	1.07	-0.01
1.09	6.00	0.00	0.780	1.09	-0.00
1.11	12.00	0.00	0.831	1.11	-0.00
1.13	20.00	0.00	0.901	1.13	0.00
1.13	26.00	0.00	0.951	1.14	0.01
1.14	30.00	0.00	0.989	1.15	0.01
3.87	6.00	0.00	1.482	3.87	0.00
3.90	2.00	0.00	1.383	3.90	-0.00
3.90	12.00	0.00	1.651	3.90	-0.00
3.93	20.00	0.00	1.877	3.93	-0.00
3.94	26.00	0.00	2.050	3.95	0.00
3.96	30.00	0.00	2.172	3.96	0.01
6.70	6.00	0.00	2.200	6.71	0.00
6.75	12.00	0.00	2.490	6.75	0.00
6.75	2.00	0.00	2.026	6.75	0.00
6.80	20.00	0.00	2.879	6.79	-0.00
6.80	30.00	0.00	3.361	6.79	-0.01
6.84	26.00	0.00	3.183	6.84	0.00

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc \* (V + Voffset) \* (1.0 + A \* T + B \* T<sup>2</sup> + C \* T<sup>3</sup>) \* Oxsol(T,S) \* exp(E \* P / K)

Residual (ml/l) = instrument oxygen - bath oxygen

