

This application note applies to the following Biospherical Instruments PAR light sensors, which all have a built-in log amplifier:

- QSP-200L and QCP-200L - no longer in production
- QSP-2300L, QSP-2350L, QCP-2300L, QCP-2300L-HP, and MCP-2300 - current production

These PAR sensors are compatible with the following Sea-Bird CTDs:

- SBE *9plus*
- SBE 16 or 19 – These PAR sensors may not be compatible with 6-cell housing version of these CTDs; consult Sea-Bird.
- SBE *16plus*, *16plus-IM*, or *19plus* – CTD's optional PAR connector **not** required when using one of these PAR sensors. The PAR sensor interfaces with an A/D voltage channel on the CTD.
- SBE *16plus V2*, *16plus-IM V2*, *19plus V2*, or *25plus* –The PAR sensor interfaces with an A/D voltage channel on the CTD.
- SBE 25 – CTD's PAR connector (optional on older versions) **not** used with these PAR sensors. The PAR sensor interfaces with an A/D voltage channel on the CTD.

**Note:** The CTD voltage channel for use with the PAR sensor can be single-ended or differential.

Seasoft computes PAR using the following equation:

$$\text{PAR} = [\text{multiplier} * (10^9 * 10^{(V-B)/M}) / \text{calibration constant}] + \text{offset}$$

Enter the following coefficients in the CTD configuration (.con or .xmlcon) file:

**M** = 1.0 and **B** = 0.0 (Notes 2 and 3)

**calibration constant** =  $10^5 / C_w$  (Notes 2 and 4)

**multiplier** = 1.0 for output units of  $\mu\text{Einstein}/\text{m}^2 \cdot \text{sec}$  (Note 5)

**offset** =  $-(10^4 * C_w * 10^V)$  (Note 6)

#### Notes:

1. In our Seasoft V2 suite of programs, edit the CTD configuration (.con or .xmlcon) file using the Configure Inputs menu in Seasave V7 (real-time data acquisition software) or the Configure menu in SBE Data Processing (data processing software).
2. Sea-Bird provides two calibration sheets for the PAR sensor in the CTD manual:
  - Calibration sheet generated by Biospherical, which contains Biospherical's calibration data.
  - Calibration sheet generated by Sea-Bird, which incorporates the Biospherical data and generates M, B, and calibration constant needed for entry in Sea-Bird software (saving the user from doing the math).
3. For all SBE 911*plus*, 16, *16plus*, *16plus-IM*, *16plus V2*, *16plus-IM V2*, 19, *19plus*, *19plus V2*, 25, and *25plus* CTDs, M = 1.0. For SBE 9/11 systems built before 1993 that have differential input amplifiers, M = 2; consult your SBE 9 manual or contact factory for further information. B should always be set to 0.0.
4.  $C_w$  is the *wet*  $\mu\text{Einstein}/\text{cm}^2 \cdot \text{sec}$  coefficient from the Biospherical calibration sheet. A typical value is on the order of  $9.0 \times 10^{-6}$ ; values have varied over the years.
5. The multiplier can be used to calculate irradiance in units other than  $\mu\text{Einstein}/\text{m}^2 \cdot \text{sec}$ . See Application Note 11General for multiplier values for other units.  
The multiplier can also be used to *scale* the data, to compare the *shape* of data sets taken at disparate light levels. For example, a multiplier of 10 would make a  $10 \mu\text{Einstein}/\text{m}^2 \cdot \text{sec}$  light level plot as  $100 \mu\text{Einstein}/\text{m}^2 \cdot \text{sec}$ .
6. Offset ( $\mu\text{Einstein}/\text{m}^2 \cdot \text{sec}$ ) =  $-(10^4 * C_w * 10^V)$ , where V is the *dark voltage*.

The dark voltage may be obtained from:

- Biospherical calibration certificate for your sensor (*Average Dark* volts on calibration sheet), or
- CTD PAR channel with the sensor covered (dark) -- in Seasave V7, display the *voltage output* of the PAR sensor channel.

#### Example:

If  $C_w = 9.0 \times 10^{-6}$  and Dark Voltage = 0.0021, offset =  $-(10^4 * 9.0 \times 10^{-6} * 10^{0.0021}) = -0.0904$ .

Instead of using the dark voltage to calculate the offset, you can also directly obtain the offset using the following method: Enter M, B, and Calibration constant, and set offset = 0.0 in the configuration (.con or .xmlcon) file. In Seasave V7, display the *calculated PAR output* with the sensor dark; then enter the negative of this reading as the offset in the configuration file.

## Mathematical Derivation

1. Using the sensor output in volts (V), Biospherical calculates:  
$$\text{light } (\mu\text{Einsteins}/\text{cm}^2 \cdot \text{sec}) = C_w * (10^{\text{Light Signal Voltage}} - 10^{\text{Dark Voltage}}).$$
2. Seasoft calculates:  $\text{light } (\mu\text{Einsteins}/\text{m}^2 \cdot \text{sec}) = [\text{multiplier} * 10^9 * 10^{(V-B)/M}] / \text{Calibration constant} + \text{offset}$   
where M, B, Calibration constant, multiplier, and offset are the Seasoft coefficients entered in the CTD configuration file.
3. To determine Calibration constant, let B = 0.0, M = 1.0, and multiplier = 1.0. Equating the Biospherical and Seasoft relationships:  
$$10^4 (\text{cm}^2 / \text{m}^2) * C_w * (10^{\text{Light Signal Voltage}} - 10^{\text{Dark Voltage}}) = (10^9 * 10^V) / \text{Calibration constant} + \text{offset}$$

Since  $\text{offset} = - (10^4 * C_w * 10^{\text{Dark Voltage}})$ , and  $V = \text{Light Signal Voltage}$ :  
$$\text{Calibration constant} = 10^9 / (10^4 * C_w) = 10^5 / C_w$$

### **Example:**

If Wet calibration factor =  $9.0 \times 10^{-6} \mu\text{Einsteins}/\text{cm}^2 \cdot \text{sec}$ , then  $C = 1.1111 \times 10^{-10}$  (for entry into configuration file).

### **Notes:**

- See Application Note 11S for integrating a Surface PAR sensor with the SBE 11*plus* Deck Unit (used with the SBE 9*plus* CTD).
- See Application Note 47 for integrating a Surface PAR sensor with the SBE 33 or 36 Deck Unit (used with the SBE 16, 16*plus*, 16*plus* V2, 19, 19*plus*, 19*plus* V2, 25, or 25*plus* CTD).

## Application Note Revision History

Date	Description
	Initial release.
July 2001	Add information about offset term, and clarify units.
October 2004	<ul style="list-style-type: none"> <li>• Update with new (2003) Biospherical PAR sensor part numbers.</li> <li>• Expand / rewrite / reorganize.</li> </ul>
June 2005	Provide output in microEinsteins/m <sup>2</sup> sec, and refer to Application Note 11 General for conversion to other units.
May 2007	Incorporate Seasave V7, and eliminate discussion of Seasoft-DOS.
March 2008	<ul style="list-style-type: none"> <li>• Update to include V2 SeaCATs (16<i>plus</i> V2, 16<i>plus</i>-IM V2, 19<i>plus</i> V2).</li> <li>• Remove ‘amps’ in description of calibration coefficients and units on Biospherical’s calibration sheet, because Biospherical changed their calibration sheet.</li> </ul>
February 2010	<ul style="list-style-type: none"> <li>• Add information on QSP-2350L (same functioning as QSP-2300L, just different connector) and QCP-2300L-HP (same functioning as QCP-2300L, just higher depth range), which are on current price list</li> <li>• Change Seasoft-Win32 to Seasoft V2.</li> <li>• Add information on .xmlcon files.</li> <li>• Update address.</li> </ul>
October 2012	Update to include SBE 25 <i>plus</i> .
August 2016	Update example / typical values of Cw and dark voltage to better reflect current Biospherical calibration sheets.