

## CALIBRATION SHEET

Instrument WXTPTU  
Serial number N5020199  
Manufacturer Vaisala Oyj, Finland  
Test date 25th January 2018

*Installed in Port Vaisala  
9 March 2018  
AMS S/N C362001*

This test report certifies that the instrument was thoroughly tested and inspected, and found to meet its published test limits when it was shipped from Vaisala.

### Calibration results

| Test phase of calibration process | Reference value | Observed value | Error*   | Uncertainty** |
|-----------------------------------|-----------------|----------------|----------|---------------|
| Pressure                          | 1078.0 hPa      | 1078.0 hPa     | 0.0 hPa  | ± 0.4 hPa     |
| Pressure                          | 896.3 hPa       | 896.4 hPa      | 0.1 hPa  | ± 0.4 hPa     |
| Pressure                          | 795.7 hPa       | 795.7 hPa      | 0.0 hPa  | ± 0.4 hPa     |
| Pressure                          | 596.2 hPa       | 596.1 hPa      | -0.1 hPa | ± 0.4 hPa     |
| Temperature                       | 59.7 °C         | 59.7 °C        | 0.0 °C   | ± 0.2 °C      |
| Temperature                       | 24.8 °C         | 24.8 °C        | 0.0 °C   | ± 0.2 °C      |
| Temperature                       | -5.7 °C         | -5.7 °C        | 0.0 °C   | ± 0.2 °C      |
| Temperature                       | -32.7 °C        | -32.7 °C       | 0.0 °C   | ± 0.2 °C      |
| Temperature                       | -51.9 °C        | -51.9 °C       | 0.0 °C   | ± 0.2 °C      |
| Relative humidity                 | 29.5 %RH        | 29.5 %RH       | 0.0 %RH  | ± 2 %RH       |
| Relative humidity                 | 57.5 %RH        | 57.5 %RH       | 0.0 %RH  | ± 2 %RH       |
| Relative humidity                 | 91.0 %RH        | 91.0 %RH       | 0.0 %RH  | ± 3 %RH       |

\*The test points for error values are polynomial fitting curve fitting points.

\*\*The calibration uncertainty given at 95 % confidence level, k = 2

### Traceability

The working standards for pressure and temperature are calibrated at Vaisala Measurement Standards Laboratory (MSL) by using MSL working standards traceable to National Institute of Standards and Technology (NIST, USA). The relative humidity values are calculated from measured temperature and dew-point temperature values. The dew-point working standards are traceable to the Finnish National Humidity Laboratory (MIKES).

Signature

*Tommy Wherry*  
\_\_\_\_\_  
Technician

This report shall not be reproduced except in full, without the written approval of Vaisala.

Doc218938-A