



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Woods Hole Oceanographic Institution
Job Number 71602
Date 11/21/2012
Technician AL

Serial Number 4560880-0340

Low Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes

High Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings. *The unit is not designed for high pressure applications.

